

# United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/699,045	10/31/2003	Daniel J. Smith	GYN-5012	1201	
27777	7590 09/22/2006		EXAMINER		
PHILIP S. JOHNSON			PRONE, CHRISTOPHER D		
JOHNSON & JOHNSON ONE JOHNSON & JOHNSON PLAZA			ART UNIT	PAPER NUMBER	
	SWICK, NJ 08933-7003		3738		
	•		DATE MAILED: 09/22/2006		

Please find below and/or attached an Office communication concerning this application or proceeding.

				ME				
	-	Application No.	Applicant(s)					
Office Action Summary		10/699,045	SMITH ET AL.					
		Examiner	Art Unit					
		Christopher D. Prone	3738					
Period fo	The MAILING DATE of this communication app or Reply	ears on the cover sheet with the	correspondence addres	SS				
WHIC - Exter after - If NO - Failu Any r	ORTENED STATUTORY PERIOD FOR REPLY CHEVER IS LONGER, FROM THE MAILING DAY IN THE MAILING	ATE OF THIS COMMUNICATIO 36(a). In no event, however, may a reply be to vill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONI	N. mely filed not the mailing date of this commuED (35 U.S.C. § 133).					
Status								
1)⊠	Responsive to communication(s) filed on 19 Ju	<u>ıly 2006</u> .						
,	This action is <b>FINAL</b> . 2b) This action is non-final.							
3)	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is							
	closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.							
Dispositi	on of Claims			,				
5)□ 6)⊠ 7)□	Claim(s) <u>1-20</u> is/are pending in the application.  4a) Of the above claim(s) <u>18-20</u> is/are withdraw Claim(s) is/are allowed.  Claim(s) <u>1-17</u> is/are rejected.  Claim(s) is/are objected to.  Claim(s) are subject to restriction and/or	n from consideration.						
Applicati	on Papers							
10)	The specification is objected to by the Examine The drawing(s) filed on is/are: a) access Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct The oath or declaration is objected to by the Ex	epted or b) objected to by the drawing(s) be held in abeyance. So ion is required if the drawing(s) is ol	ee 37 CFR 1.85(a). bjected to. See 37 CFR 1					
Priority u	under 35 U.S.C. § 119							
<ul> <li>12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).</li> <li>a) All b) Some * c) None of:</li> <li>1. Certified copies of the priority documents have been received.</li> <li>2. Certified copies of the priority documents have been received in Application No</li> <li>3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).</li> <li>* See the attached detailed Office action for a list of the certified copies not received.</li> </ul>								
2)  Notice  No	t(s) se of References Cited (PTO-892) se of Draftsperson's Patent Drawing Review (PTO-948) mation Disclosure Statement(s) (PTO/SB/08) sr No(s)/Mail Date	4) Interview Summar Paper No(s)/Mail D 5) Notice of Informal 6) Other:	Date					

#### **DETAILED ACTION**

## Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 3-14, 16, and 17 are rejected under 35 U.S.C. 102(b) as being anticipated by United States Patent 5,951,525 Thorne et al.

Thorne discloses the same invention comprising a stem portion (30) having a recess (32) (shown best in figure 2) along its length being shaped to receive a surgical needle (6:36-52), having a c-shaped cross-section at its distal end (32) and first and second extension portions comprising stationary portions (the left and right portions of (80)) coupled to the stem, movable portions (100) and (110) bendably/pivotably coupled to the stationary portions, and gripping elements (121) and (141).

In regards to claims 7 and 8, Thorne discloses extension portions that extend from the stem portion at angles ranging from 45-180 degrees. The range of angles the extension portions extend from the stem is shown best in figures 1, 3, and 4.

### Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and

Application/Control Number: 10/699,045

Art Unit: 3738

the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 2 and 15 rejected under 35 U.S.C. 103(a) as being unpatentable over United States Patent 5,951,525 Thorne et al.

Thorne discloses the invention substantially as claimed being a guide for use in medical procedures as disclosed above. However Thorne fails to disclose the exact distance between the extension portions and the stem. It would have been an obvious matter of design choice to make the distance in the first position range from 55-65 mm and the distance in the second position range from 65-75 mm in order to accommodate larger or obese patients. Such a change involving a mere change in size is recognized as being within the level of ordinary skill in the art.

#### Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Application/Control Number: 10/699,045

Art Unit: 3738

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

> Christopher D Prone Examiner Art Unit 3738

CORRINE McDERMOTT SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700

Page 4